



HAL
open science

Data on near infrared polarization spectroscopy measurements to evaluate the potential of the Mueller matrix elements in characterization of turbid liquid samples

Arnaud Ducanhez, Daphné Heran, R. Bendoula

► To cite this version:

Arnaud Ducanhez, Daphné Heran, R. Bendoula. Data on near infrared polarization spectroscopy measurements to evaluate the potential of the Mueller matrix elements in characterization of turbid liquid samples. *Data in Brief*, 2019, 23, pp.103756. 10.1016/j.dib.2019.103756 . hal-02609243

HAL Id: hal-02609243

<https://hal.inrae.fr/hal-02609243>

Submitted on 16 May 2020

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



ELSEVIER

Contents lists available at ScienceDirect

Data in brief

journal homepage: www.elsevier.com/locate/dib



Data Article

Data on near infrared polarization spectroscopy measurements to evaluate the potential of the Mueller matrix elements in characterization of turbid liquid samples



A. Ducanhez^{*}, D. Heran, R. Bendoula

Montpellier SupAgro, Irstea, Univ. Montpellier, UMR ITAP, 34060 Montpellier, France

ARTICLE INFO

Article history:

Received 3 December 2018

Received in revised form 10 January 2019

Accepted 7 February 2019

Available online 7 March 2019

ABSTRACT

In this article, a set of 50 turbid liquid samples with different levels of absorption and scattering properties were prepared and measured at various orientations of polarizers and analyzers to obtain the 16 elements of the complete Muller matrix. Partial Least Square (PLS) was used to build calibration models in order to assess the potential of polarization spectroscopy through the elements of Muller matrix to predict chemical and physical parameters.

© 2019 The Authors. Published by Elsevier Inc. This is an open access article under the CC BY license (<http://creativecommons.org/licenses/by/4.0/>).

^{*} Corresponding author.

E-mail address: arnaud.ducanhez@supagro.fr (A. Ducanhez).

Specifications table [Please fill in right-hand column of the table below.]

Subject area	Optical Diffuse Spectroscopy, Chemistry, Scattering media.
More specific subject area	Polarization Light Spectroscopy
Type of data	Table, Figure, .mat file
How data was acquired	Polarized Light System with PSG (polarization state generator) and PSA (polarization state analyzer) coupled with Visible and Near Infrared Spectrometer (Zeiss MSS 1)
Data format	Raw and analyzed data
Experimental factors	50 turbid liquid samples that contain different levels of absorption and scattering properties were measured in different polarization states in order to obtain the elements of Muller matrix
Experimental features	Polarization light Spectroscopy coupled with chemometric analysis was used to assess the potential of Muller matrix to predict and to discriminate the concentration in absorbers and scatterers of turbid liquid samples
Data source location	IRSTEA, Montpellier, France (43° 36' 38.768" N 3° 52' 36.178" E)
Data accessibility	Data is available with this publication
Related research article	W.S. Bickel and W.M. Bailey, "Stokes vectors, Mueller matrices, and polarized light scattering," <i>Am. J.Phys.</i> 53, 468–478 (1985).

Value of the data

- These data describe the complete Muller matrix in visible (VIS) and near-infrared (NIR) wavelength range for different model turbid media.
- These data establish a link between physical and/or chemical properties of various turbid liquid samples and VIS/NIR spectra in diffuse reflectance.
- VIS/NIR spectroscopy in polarized light and multivariate analysis are able to predict physical and chemical properties of turbid media.
- The experimental data of turbid media coupled with VIS/NIR spectra can be used as reliable database for testing and analyzing different PLS methods as multi-block regression (PO-PLS, SO-PLS ... etc.) [1].

1. Data

Several measurements on 50 turbid liquid samples with different level of concentration in scatterers and absorbers were made (Fig. 1). In parallel, diffuse reflectance spectra in different states of polarized light in visible and near-infrared wavelength range were measured with Polarized Light System (Fig. 2). From these spectral raw data, linear combination of these different polarization states were calculated to obtain the complete Mueller matrices (Fig. 3) for each 50 turbid liquid samples corresponding to the processed data set. The details of all databases are given in part 3. With a Partial Least Square (PLS) algorithm, an example of predicting models were obtained for the concentration of absorbers (Fig. 4a) and scatterers (Fig. 4b) for element M_{22} of the Muller matrix.

2. Experimental design, materials, and methods

2.1. Liquid phantom

Turbid liquid samples were composed of methylene blue (MB), Intralipid 20% solution (IL) and distilled water used as absorbing, scattering and dilution material respectively [2]. IL is an intravenous fat emulsion that contains fat globules which act as scattering particles. MB is a water-soluble non-scattering dye that presents two peaks absorption at 668 nm and 609 nm due to monomer and dimer forms in aqueous solutions [3]. In contrast, the absorption by water and IL is minimal in the wavelength range. Hence, MB and IL are well-adapted in the considered wavelength range to interpret and to discriminate the effects of scattering and absorption on the measured polarized reflectance spectra.

2.2. Samples preparation

A set of 50 turbid liquid samples were prepared by mixing in different ratios 10 concentration of the absorber (0, 1, 2, 4, 5, 8, 12, 16, 20 and 32 mL of a 400 μ M stock solution) relating to row number with 5

		Scattering Level				
		A	B	C	D	E
Absorption Level	1					
	2					
	3					
	4					
	5					
	6					
	7					
	8					
	9					
	10					

Fig. 1. Set of 50 turbid liquid samples obtained by mixing MB (absorber), IL (scatterer) and distilled water (dilution agent). For rows, MB concentrations increases from top to bottom (0, 4, 8, 16, 20, 32, 48, 64, 80 and 128 μM) and for columns, IL concentrations increases from left to right (0.227, 0.454, 0.908, 1.816 and 3.682%).

concentrations of the scattered (1, 2, 4, 8 and 16 mL of the IL 20% solution) relating to column letter [Fig. 1]. According to this experience plan, the liquid samples have MB concentration of 0, 4, 8, 16, 20, 32, 48, 64, 80 and 128 μM respectively with increasing row number and a scattering particles concentration of 0.227, 0.454, 0.908, 1.816 and 3.682% (v/v) respectively with increasing column letter. Each sample was conditioned in a beaker and distilled water was added to obtain 100 mL for all liquid samples. At last, each sample was rigorously mixed to homogenize the aqueous solutions before the spectral measurements.

2.3. Polarization spectroscopy system

The experimental setup for studying the diffuse reflectance of polarized light is presented in Fig. 2. This system was the same Polarized Light Spectroscopy system (PoLiS) that we used in previous study [4]. The PoLiS system integrates a polarization state generator (PSG) and polarization state analyzer (PSA). Both the PSG and PSA consisted of a rotating broad-band (400–800 nm) linear polarizer (LP) (NT52-557, Edmunds Optics) and a rotating quarter-wave plate (QWP) (AQWP05M – 600, Thorlabs) in order to generate and to select the various polarization states that are needed. Spectral data were collected with a spectrometer (MMS1, Zeiss) in the 350–1100 nm wavelength range at 3 nm intervals.




Fig. 2. Schematic representation of Polarized Light Spectroscopy (PoLiS) system in reflexion [4] (PSG = polarization state generator, PSA = polarization state analyzer, LP = linear polarizer, QWP = quarter wave plate and AL = aspheric lens).

M_{11} : (OO)	M_{12} : (HO-VO)/2	M_{13} : (PO-MO)/2	M_{14} : (LO-RO)/2	<div style="display: flex; flex-direction: column; align-items: center;"> <div style="width: 20px; height: 20px; background-color: #fce4d6; border: 1px solid black; margin-bottom: 5px;"></div> <div style="width: 20px; height: 20px; background-color: #e2efda; border: 1px solid black; margin-bottom: 5px;"></div> <div style="width: 20px; height: 20px; background-color: #e0f2f1; border: 1px solid black;"></div> </div> <div style="margin-top: 5px;"> <i>Polarization properties</i> <i>Depolarization properties</i> <i>Birefringence properties</i> </div>
M_{21} : (OH-OV)/2	M_{22} : (HH+VV)/4 - (HV+VH)/4	M_{23} : (PH+MV)/4 - (PV+MH)/4	M_{24} : (LH+RV)/4 - (LV+RH)/4	
M_{31} : (OP-OM)/2	M_{32} : (HP+VM)/4 - (HM+VP)/4	M_{33} : (PP+MM)/4 - (PM+MP)/4	M_{34} : (LP+RM)/4 - (LM+RP)/4	
M_{41} : (OL-OR)/2	M_{42} : (HL+VR)/4 - (HR+VL)/4	M_{43} : (PL+MR)/4 - (PR+ML)/4	M_{44} : (LL+RR)/4 - (RL+LR)/4	

Fig. 3. This table details how to obtain the different elements of the Muller matrix from combinations of different states of polarization [5]. A two letter combination corresponds to one measurement. For example, the combination (VP) means that the PSG is adjusted to obtain linear polarization along the vertical axis (y axis) for incoming light and PSA is adjusted to recover linear polarization with a + 45° offset for reflected light.




Fig. 4. Calibration model for the concentrations of methylen blue absorbers (c_{abs}) in μM (a) and the concentrations of intralipid scatters (c_{dif}) in % (b) of element M_{22} of the Muller matrix.

2.4. Mueller matrix

The Mueller matrix provides the most general and complete description of the response of a medium to excitation by polarized light in either reflection or transmission configurations. Fig. 3 lists the necessary measurements and combinations in order to determinate each matrix element [5]. To obtain the 16 elements of the complete Mueller matrix, it was necessary to measure 7 configurations respectively for both the PSG and PSA:

- O: unpolarized
- H: linearly polarized along the horizontal axis (x axis),
- V: linearly polarized along the vertical axis (y axis),
- P: linearly polarized with a $+45^\circ$ offset,
- M: linearly polarized with a -45° offset,
- L: left-handed circularly polarized,
- R: right-handed circularly polarized,

From these 49 configurations in total corresponding to 49 reflectance spectral measurements (7 for incoming polarized light x 7 for reflected polarized light), each matrix element was calculated by the corresponding linear combinations. For example, the M_{12} element is obtained by measuring the total reflected intensity for an incident light linearly polarized along the horizontal axis and subtracting from this the total reflected intensity for an incident light linearly polarized along the vertical axis. Finally, except for the first element, each other element of the Muller matrix corresponding to physical properties of the medium considered such as polarization, depolarization or birefringence.

2.5. Spectral acquisition

For each turbid liquid samples, diffuse reflectance spectra of polarized light was measured for each of the 49 configurations from the both PSG and PSA. Moreover, for each of these 49 configurations, white diffuse standard (Spectralon®, SRS-99-010. Labsphere) was measured in order to standardize spectra from non-uniformities of all components of the PoLiS system.

2.6. PLS algorithm

All computations and multivariate data analysis were performed with Matlab software v. R2015b (The Mathworks Inc., Natick, MA, USA). A Partial Least Square (PLS) algorithm was used to model the physical and chemical parameters of the turbid liquid media. A general PLS model was built using the whole calibration set (two third of the sample) and a predicting set (one third of the sample). The number of Latent Variables (LV) was determined by comparing performances by leave-one-out cross validation method. Two basic statistical parameters including the determination coefficient (R^2) and the standard error of cross-validation (SECV) were calculated. These parameters were used to assess the performance of each calibration model for predicting absorbers and scatterers concentration. An example of predicting models for element M_{22} of the Muller matrix is presented for the concentration of absorbers (Fig. 4a) and scatterers (Fig. 4b).

3. Details on the data available

In this publication, several database matrices in matlab file are available:

- "raw_spectral_data": a zip file that contains 50 matlab files corresponding to the 50 turbid liquid samples. Each matlab file contains 49 raw spectral data in diffuse reflectance from the 49 possible configurations (OO, OR, ..., VV.) through the different polarization states at the input and output of the PoLiS system necessary to create the complete Mueller matrix.

- "cabs": matrix with 50 rows and 1 column corresponding to the concentrations of absorbers in μM for the 50 turbid liquid samples with this structuring:

$$cabs = \begin{bmatrix} A1 \\ B1 \\ C1 \\ D1 \\ E1 \\ A2 \\ B2 \\ \vdots \\ E10 \end{bmatrix} = \begin{bmatrix} 0 \\ 0 \\ 0 \\ 0 \\ 0 \\ 4 \\ 4 \\ \vdots \\ 128 \end{bmatrix}$$

Note that the concentration of absorbers increases with the number.

- “*cdif*”: matrix with 50 rows and 1 column corresponding to the concentrations of scatterers in % for the 50 turbid liquid samples with this structuring:

$$cdif = \begin{bmatrix} A1 \\ B1 \\ C1 \\ D1 \\ E1 \\ A2 \\ B2 \\ \vdots \\ E10 \end{bmatrix} = \begin{bmatrix} 0.227 \\ 0.454 \\ 0.908 \\ 1.816 \\ 3.682 \\ 0.227 \\ 0.454 \\ \vdots \\ 3.682 \end{bmatrix}$$

Note that the concentration of scatterers increases with the letter.

- “*m11, m22 ... m14*”: 16 matrices corresponding to the 16 elements (4×4) of the Mueller matrix for each of the 50 liquid phantom samples. Each matrices consists of 50 rows (50 liquid phantom samples) and 256 columns (256 wavelength of spectrometer in the 303–1146 nm wavelength range at 3.3 nm intervals).

For example, *m11* matrix integrates the first element of each of Muller’s 50 matrices and is therefore structured as follows:

$$m11 = \begin{bmatrix} [A1; \lambda_{303}] & [A1; \lambda_{306}] & \dots & [A1; \lambda_{1146}] \\ [B1; \lambda_{303}] & [B1; \lambda_{306}] & \dots & [B1; \lambda_{1146}] \\ [C1; \lambda_{303}] & [C1; \lambda_{306}] & \dots & [C1; \lambda_{1146}] \\ [D1; \lambda_{303}] & [D1; \lambda_{306}] & \dots & [D1; \lambda_{1146}] \\ [E1; \lambda_{303}] & [E1; \lambda_{306}] & \dots & [E1; \lambda_{1146}] \\ [A2; \lambda_{303}] & [A2; \lambda_{306}] & \dots & [A2; \lambda_{1146}] \\ [B2; \lambda_{303}] & [B2; \lambda_{306}] & \dots & [B2; \lambda_{1146}] \\ \vdots & \vdots & \dots & \vdots \\ [E10; \lambda_{303}] & [E10; \lambda_{306}] & \dots & [E10; \lambda_{1146}] \end{bmatrix}$$

- “*r11, r22 ... r14*”: 16 matrices corresponding to “*m11, m22 ... m14*” matrices where each element was divided by the corresponding spectral reference in order to normalize the spectral measurements.

- “*lbd*”: the 256 wavelengths of spectrometer.

Acknowledgments

This dataset was funded by French National Research Agency with the reference ANR-16-CE04-0010 (OptiPAG project).

Transparency document

Transparency document associated with this article can be found in the online version at <https://doi.org/10.1016/j.dib.2019.103756>

Appendix A. Supplementary data

Supplementary data to this article can be found online at <https://doi.org/10.1016/j.dib.2019.103756>.

References

- [1] A. Biancolillo, *Method Development in the Area of Multi-Block Analysis Focused on Food Analysis*, PhD Thesis, Department of Food Science, Faculty of Science, University of Copenhagen, 2016.

- [2] B. Aernouts, E. Zamora-Rojas, R. Van Beers, R. Watté, L. Wang, M. Tsuta, J. Lammertyn, W. Saeys, Supercontinuum laser based optical characterization of Intralipid® phantoms in the 500-2250 nm range, *Opt. Express* 21 (2013) 32450–32467.
- [3] E. Genina, A. Bashkatov, V. Tuchin, Methylene blue diffusion in skin tissue, *Proc. SPIE* 5486, ALT'03 International Conference on Advanced Laser Technologies: Biomedical Optics, 29 July 2004. <https://doi.org/10.1117/12.572047>.
- [4] A. Gobrecht, R. Bendoula, J.-M. Roger, V. Bellon-Maurel, Combining linear polarization spectroscopy and the representative layer theory to measure the Beer-Lambert law absorbance of highly scattering materials, *Anal. Chim. Acta* 853 (2015) 486–494.
- [5] A. Hielscher, A. Eick, J. Mourant, D. Shen, J. Freyer, I. Bigio, Diffuse backscattering Mueller matrices of highly scattering media, *Opt. Express* 1 (1997) 441–453.