



HAL
open science

Determination of reaction orders for ascorbic acid degradation during sterilization using a new experimental device: The thermoresistometer Mastia®

Nizar Al Fata, Stéphane Georgé, Stéphane André, Catherine M.G.C. Renard

► **To cite this version:**

Nizar Al Fata, Stéphane Georgé, Stéphane André, Catherine M.G.C. Renard. Determination of reaction orders for ascorbic acid degradation during sterilization using a new experimental device: The thermoresistometer Mastia®. *LWT - Food Science and Technology*, 2017, 85, pp.487-492. 10.1016/j.lwt.2016.08.043 . hal-02622313

HAL Id: hal-02622313

<https://hal.inrae.fr/hal-02622313v1>

Submitted on 26 May 2020

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Determination of reaction orders for ascorbic acid degradation during sterilization using a new experimental device: The thermoresistometer Mastia[®]

Nizar Al Fata^{a, b, c, d, *}, Stéphane Georgé^a, Stéphane André^b, Catherine M.G.C. Renard^{c, d}

^a CTCPA, unité qualité nutritionnelle, Site Agroparc, 449 Clément Ader, BP 21203, 84911 Avignon Cedex 9, France

^b CTCPA, unité de microbiologie EMaiRIT'S, Site Agroparc, 449 Clément Ader, BP 21203, 84911 Avignon Cedex 9, France

^c INRA, UMR408 Sécurité et Qualité des Produits d'Origine Végétale, Domaine Saint Paul, Site Agroparc, F-84000 Avignon, France

^d Avignon University, UMR408 Sécurité et Qualité des Produits d'Origine Végétale, Domaine Saint Paul, Site Agroparc, F-84000 Avignon, France

A B S T R A C T

Sterilization is the most relevant technique to guaranty safety of shelf-stable highly hydrated food products at ambient temperature. However, these thermal processes are often optimised to microbiological aspect and consequently are not without consequences on food quality. Ascorbic acid's degradation at sterilization temperature (>100 °C) is poorly documented. Using a new instrumented reactor (thermoresistometer Mastia[®]), the behavior of ascorbic acid was studied in food model solution at pH = 3.5, in aerobic and anaerobic conditions from 95 °C to 125 °C. In anaerobic conditions, degradation was slow and followed first order kinetics ($k_{125^{\circ}\text{C}} = 2.88 \cdot 10^{-3} \text{ min}^{-1}$, $E_a = 96 \text{ kJ/mol}$). In aerobic conditions, reactions were faster with total elimination of ascorbic acid in about 4 h, and the data were best fitted using 0.5-order, while reaction speed varied little with temperature.

Keywords:

Vitamin C
 Canning
 Modelling
 Kinetics
 Nutritional qualities

1. Introduction

Ascorbic acid is often used as an indicator molecule in food processing because it is highly reactive and very sensitive to physico-chemical conditions in the medium. Indeed, ascorbic acid degradation is influenced by the presence of oxygen, temperature (Lin & Agalloco, 1979; Torregrosa, Esteve, Frígola, & Cortés, 2006), light intensity (Bosch et al., 2013; Sapei & Hwa, 2014), pH, water activity (Berlinet, Brat, Brillouet, & Ducruet, 2006; Mercali, Jaeschke, Tessaro, & Marczak, 2012), °Brix (Al-Zubaidy & Khalil, 2007), the presence of metallic ions such as Fe²⁺ or Cu²⁺ (Bosch et al., 2013; Lin & Agalloco, 1979; Soares & Hotchkiss, 1999), presence of sugars (Hsu, Tsai, Fu, & Wu, 2012), amino acids, sugars, lipids and enzymes (Mercali, Schwartz, Marczak, Tessaro, & Sastry, 2014), or microorganisms (Soares & Hotchkiss, 1999). Storage conditions (packaging, temperatures of storage, etc ...) also have an

effect on the shelf-life of ascorbic acid (Mercali et al., 2012; Torregrosa et al., 2006).

Thermal treatment is the most popular and efficient method to inactivate microorganisms to stabilize canned vegetables (Blasco, Esteve, Frígola, & Rodrigo, 2004; Peng et al., 2012; Zimmermann, Longhi, Schaffner, & Aragão, 2014). There is an increased demand of nutritive foods, thereby, there were many attempts to maximise nutrients retention during industrial processes as well as during transport and storage (Sapei & Hwa, 2014). To be efficient, a thermal treatment must ensure that the product has been exposed at sufficiently high temperature for a sufficient time to destroy an appropriate number of targeted microorganisms (Zimmermann et al., 2014). Nevertheless, thermal treatments can generate undesirable chemical reactions like vitamin degradation, which can lead to products having less nutritive interest. In the literature, only few studies are focussed on the degradation of ascorbic acid at sterilization temperature (Blasco et al., 2004; Esteve, Frígola, Martorell, & Rodrigo, 1998; Oey, Verlinde, Hendrickx, & Van Loey, 2006; Van den Broeck, Ludikhuyze, Weemaes, Van Loey, & Hendrickx, 1998). It is generally assumed that the degradation of ascorbic acid follows first-order kinetics in food systems (Jiang, Zheng, & Lu, 2014), but it can follow second-order kinetics if oxygen is present in a limited

* Corresponding author. CTCPA, unité qualité nutritionnelle, Site Agroparc, 449 Clément Ader, BP 21203, 84911 Avignon Cedex 9, France.

E-mail addresses: nalfata@ctcpa.org (N. Al Fata), sgeorge@ctcpa.org (S. Georgé), sandre@ctcpa.org (S. André), catherine.renard@avignon.inra.fr (C.M.G.C. Renard).

amount in the media (Lin & Agalloco, 1979). Ascorbic acid degradation at sterilization temperatures is poorly known, probably because of the difficulty to work at above 100 °C (Blasco et al., 2004; Esteve et al., 1998; Oey et al., 2006; Van den Broeck et al., 1998). First order kinetics were reported by Van den Broeck et al. (1998) in orange and tomato juices in anaerobic conditions between 120 and 150 °C, by Esteve et al. (1998) in green asparagus extracts both in aerobic and anaerobic conditions between 110 and 140 °C, and by Blasco et al. (2004) in mushroom extracts both in aerobic and anaerobic conditions between 110 and 140 °C. In the last two cases, degradation was faster under aerobic condition than under anaerobic conditions. Oey et al. (2006) state that ascorbic acid degradation follows first order kinetics in a case where oxygen is limited between 110 and 120 °C. However they report ascorbic acid degradation profiles which display two steps: a first, very rapid step, corresponding to the aerobic pathway where dissolved oxygen is involved, and a second, slower step when oxygen has been consumed and the anaerobic pathway occurs.

Our goal was to study the degradation of ascorbic acid depending on two factors, the availability of oxygen and temperature, in conditions relevant for the canning industry (temperature > 100 °C, pressure > 0.1 MPa). The thermoresistometer Mastia[®] is used in microbiology in order to determine the heat resistances of microorganisms, both in model solution and in various matrices, during isothermal or non-isothermal treatments and for more than 20 years (Condón, Arrizubieta, & Sala, 1993; Garza, Teixidó, Sanchis, Viñas, & Condón, 1994; Palop, Marco, Raso, Sala, & Condón, 1997; Palop, Raso, Pagàn, Condón, & Sala, 1999; Palop, Sala, & Condón, 1997; Raso, Palop, Bayarte, Condón, & Sala, 1995). The accuracy of the thermoresistometer allows determining *D*-values close to 10⁻³ min (Condón et al., 1993), especially due to its fast homogenization. It works at temperature higher than 100 °C, under high pressure and at modified gas composition and therefore was appropriate to our goal.

2. Materials & methods

2.1. Reagents

All the reagents used were purchased from Fisher Chemical (Illkirch, France) and are analytical grade or better.

2.2. Preparation of ascorbic acid solution

A 5.0000 g accurately weighed portion of L-ascorbic acid was dissolved in 25 mL of 23.1 mL/L acetic acid solution (pH = 2.6) in a 25 mL volumetric flask. The mixture was then stirred until complete dissolution of ascorbic acid. A volume of 1.8 mL of this solution was injected inside the vessel of the thermoresistometer, in order to have an initial concentration of ascorbic acid in the heating media at 90 mg/100 mL.

2.3. Model solution

Citrate-phosphate McIlvaine buffer (pH = 3.5) was used as heating medium. McIlvaine buffer was prepared using a 0.1 mol/L monohydrate citric acid solution and a 0.2 mol/L disodium phosphate solution (McIlvaine, 1921). The buffer was stored at 4 °C until used.

2.4. High-performance liquid chromatography

Analysis of ascorbic acid was carried out using a 1260 Agilent infinity LC chromatograph with a 1290 Agilent diode-array detector. A 150 × 3.0 mm Synergi Polar RP-C18 column (Phenomenex)

was used with a precolumn and both have a particle size of 4 μm. The mobile phase was a 23.1 mL/L acetic acid solution as eluent A and acetonitrile as eluent B. The flow rate was fixed at 0.8 mL min⁻¹, the column temperature was set at 30 °C and the injection volume was 5 μL. A gradient program was performed as follows: the initial conditions were 96% A/4% B; 0–5 min, 96% A/4% B; 5–10 min, 80% A/20% B; 10–15 min, 96% A/4% B. The data acquisition was assessed at 245 nm, corresponding to the maximum absorbance of ascorbic acid in UV-visible spectrometry (Jaffe, 1984). Quantification of ascorbic acid was carried out by external standard method with a calibration curve, which is the mean of five calibration curves established by five standard solutions of ascorbic acid injected in HPLC in the same conditions.

2.5. Determination of heat resistance by the thermoresistometer Mastia[®]

Thermal treatments were carried out in a thermoresistometer TR-SC Mastia[®] (Fig. 1), the operation system of which is described by Conesa et al. (Conesa, Andreu, Fernández, Esnoz, & Palop, 2009). It works in a temperature range from 20 °C to 150 °C under pressure, using different gases like air or nitrogen, having a constant and regulated stirring, sampling or injecting at any time during the heat treatment, even under pressure, without perturbation of the experimental media, working at large volumes (maximum 400 mL), and simulating isothermal and non-isothermal heat treatments.

The thermoresistometer TR-SC possesses a 2 kW heating electric element (2) and a cooling ring (1) in which cold water (2 °C) provided by the cooling system (A) is circulating. This device is implemented with a programmable logic control (PLC) (B), which powers the heating electric element and the cooling system. The programmable logic control is connected to a tactile screen (C) for communication with to the PLC, and a computer (D) provided with a software able to program or/and register temperature profiles (Conesa et al., 2009). The working media is put in a classical stainless vessel or a teflonated stainless vessel (8.5 × 12 cm outer diameter) respectively for microbiological or chemical studies (E), which is screwed on the thermoresistometer cap with an O-ring (F). This cap has a stirring shaft with a propeller (3) which is powered by the stirring motor (I), and eight ports with screw cap; one is holding the pressure source (G), another is the injection port and




Fig. 1. Diagram of the Thermoresistometer Mastia[®] set up. (A) Cooling water. (B) Programmable Logic Controller (PLC). (C) Tactile screen. (D) Computer. (E) Vessel. (F) Cap. (G) Pressure source (air or N₂). (H) Hamilton-type syringe. (I) Stirring motor. (J) Proportional Integral Derivative (PID). (1) Cooling ring. (2) Electric heating element. (3) Propeller. (4) Sampling tube. (5) Thermocouple. (6) Manometer. (7) Solenoid sampling valve. (8) Solenoid valve for the cooling system redrawn from Conesa et al. (2009).

Comment citer ce document :

Al Fata, N. (Auteur de correspondance), Georgé, S., André, S., Renard, C.M.G.C. (2017).
 Determination of reaction orders for ascorbic acid degradation during sterilization using a new experimental device: The thermoresistometer Mastia[®]. LWT - Food Science and Technology, 85, 487-492.
 DOI : 10.1016/lwt.2016.08.043

contain a gas chromatography septum, a third one holds the sampling tube (4), another holds the thermocouple (5), two ports maintain the electric element and two others maintain the cooling ring.


The vessel can be pressurized by a manometer (6) connected to the pressure source (G). Pressurization is needed to allow extraction of samples and to avoid boiling at temperatures higher than 100 °C in aqueous solutions. A specific Hamilton-type syringe (H) was used to inject samples in the vessel in order to overcome the pressure.

For all experiments, the over pressure of the vessel was set at 0.2 MPa, i.e a total pressure of 0.3 MPa. The temperature range studied is from 95 °C to 125 °C. Aerobic conditions were obtained using air as headspace gas. Anaerobic conditions were performed by removal of oxygen from the media by heating to 95 °C during 30 min and bubbling nitrogen *in situ* during 30 min at 20 °C through the sampling tube, and then using nitrogen as headspace gas. Establishment of anaerobic conditions was validated by measuring oxygen concentration after each step using a PreSens Fibox 4 device (Regensburg, Germany) coupled to a fiber optic probe with a sensor type PSt3 (limit of detection 30 Pa oxygen, 0.015 mg/L dissolved oxygen). Each thermal treatment lasted 320 min.

2.6. Kinetic models

Generally, it is assumed that ascorbic acid (AA) degradation can follow two pathways: one is the oxidoreductive (aerobic) pathway by the formation of dehydroascorbic acid (DHAA), and another is by hydrolytic (anaerobic) pathway, by direct cleavage of the lactone ring of ascorbic acid molecule (Yuan & Chen, 1998). The principal characteristics of these two mechanisms are rapidly described as follow:

2.6.1. Oxidoreductive mechanism



The formation of DHAA from AA is reversible depending on redox potential. It is assumed that the reformation of AA from DHAA is negligible unless the medium is specifically made reductive, and this is more true at high temperature (Serpen & Gökmen, 2007). DHAA is then irreversibly degraded to 3-diketogulononic acid.

2.6.2. Anaerobic pathway

This degradation pathway is akin to sugar degradation and does not require oxygen.

The global ascorbic acid degradation taking in account both mechanisms can be described by the following differential equation:

$$\frac{d[AA]}{dt} = -k_{ox} \cdot [AA]^\alpha [O_2]^\beta - k_H \cdot [AA]^\gamma \quad (1)$$

where [AA] is the ascorbic acid concentration, [O₂] the dissolved oxygen concentration, k_{ox} and k_H are the kinetic constants respectively for aerobic and anaerobic degradation, α is the partial kinetic order regarding ascorbic acid concentration for aerobic degradation, β the partial kinetic order regarding dissolved oxygen concentration and γ the partial order regarding ascorbic acid concentration for anaerobic degradation.

In anaerobic conditions with [O₂] = 0, equation (1) can be simplified to:

$$\frac{d[AA]}{dt} = -k_H \cdot [AA]^\gamma \quad (2)$$

Ascorbic acid degradation in anaerobic or aerobic conditions was fitted by using general n -order kinetics model in order to determine the partial reaction order parameter α or γ :

$$n - \text{order kinetics} : C_t = \left(C_0^{1-n} + (n-1) \cdot k(T) \cdot t \right)^{\frac{1}{1-n}} \quad (3)$$

Starting from equation (3), we linearized the equation in order to obtain a straight line curve as described by the following equation:

$$C_t^{1-n} = C_0^{1-n} + (n-1) \cdot k \cdot t \quad (4)$$

where $(n-1) \cdot k$ is the slope of the curve and C_0^{1-n} is the intercept.

The kinetic constant k is then determined by taking the inverse of the slope of the curve described by equation (4).

From these results, 0.5-order kinetics and first-order kinetics models were again used to model ascorbic acid degradation:

$$0.5 - \text{order kinetics} : C_t = C_0 - k(T) \cdot t \cdot \sqrt{C_0} + \frac{(k(T) \cdot t)^2}{4} \quad (5)$$

$$\text{first - order kinetics} : C_t = C_0 \cdot \exp(-k(T) \cdot t) \quad (6)$$

where C_t is the ascorbic acid concentration at time t , C_0 is the initial ascorbic acid concentration, k is the rate constant and n is the partial reaction order of ascorbic acid, T is the temperature and t is the time in minute.

The temperature dependence of k was fitted using Arrhenius law:

$$k(T) = A \cdot \exp\left(\frac{-E_a}{RT}\right) \quad (7)$$

where A is the frequency factor (also called pre-exponential factor), E_a is the activation energy, and R is the ideal gas constant.

2.7. Analysis of data

All experiments were repeated at least two times and each point was sampled twice; the results were reported as average.

2.8. Statistical analysis

To evaluate the fit of a model fit to the data, we analysed residual plots between the data and the model. Shapiro-Wilk test was used to determine if residues were normally distributed or not at a chosen alpha level of 0.05. The p-value was analysed to discriminate if the distribution is normal or not. When the p-value was superior to the α -level, we considered that the residues are normally distributed.

3. Results & discussion

3.1. Anaerobic conditions

In anaerobic conditions (Fig. 2) not all the ascorbic acid was degraded even after 320 min and regardless of the temperature. Moreover, after 320 min, degradation of ascorbic acid was only about 8% at 95 °C, around 18% at 105 °C, around 36% at 115 °C and a maximum of degradation around 61% at 125 °C. Fitting all curves simultaneously gave a reaction order $n = 0.83$. As analysed by Van




Fig. 2. Ascorbic acid degradation in anaerobic conditions at 95 °C (◇), 105 °C (□), 115 °C (Δ) and 125 °C (x) in Mcllvaine buffer pH = 3.5.

Boeckel, no real distinction can be made between models if the fractional conversion is less than 60–70% (Van Boeckel, 2009). Indeed, we can legitimately base the determination of order reaction only on the degradation curve at 125 °C, with a degradation of 61%, whereas it is less than 20% at the other temperatures. The reaction order determined at 125 °C was $n = 0.84$. We chose thereby to use first-order kinetics to fit the data. The R^2 was compared for $n = 0.84$ and $n = 1$ in order to confirm that there is no significant difference between the calculated value of n and the fit of $n = 1$. A good fit was obtained for all four temperatures using $n = 1$ ($R^2 > 0.99$) and the R^2 was the same for the two different n -values (Table 1). The residues were randomly dispersed in the residual plot between the data and the obtained values with first-order kinetics model at 125 °C, and the p-value was 0.91 at an α -level of 0.05, i.e. the residues are normally distributed.

The values of first-order kinetic constants are reported in Table 1. The k_H -values obtained are lower than those obtained by Van den Broeck et al. (1998), Blasco et al. (2004), Esteve et al. (1998) and Oey et al. (2006), probably due to the fact that in our case the anaerobic conditions were strict (the heating media was deaerated before each experiment and nitrogen was used as headspace gas), while in their cases, no deaeration is mentioned before heating treatment and therefore a certain amount of oxygen was present in the treated media.

There was a clear impact of temperature and the kinetic constants were used to calculate Arrhenius parameters (Fig. 3). The activation energy determined had a value of 96 kJ/mol. This value is of the same order of magnitude though slightly lower than those reported for ascorbic acid degradation in plant foods at canning temperature: under anaerobic conditions, Van den Broeck et al. (1998) have determined activation energies of 117 kJ/mol and 105 kJ/mol respectively for ascorbic degradation in orange juice and tomato juice between 120 °C and 150 °C, Esteve et al. (1998) have determined an activation energy of 140 kJ/mol for ascorbic acid degradation in green asparagus.

Table 1
 Kinetic constant as function of temperature for ascorbic acid degradation in anaerobic conditions (first-order in 10^{-4} min^{-1}) and in aerobic conditions (0.5-order in $10^{-4} \text{ mg}^{1/2} \cdot 100 \text{ mL}^{-1/2} \cdot \text{min}^{-1}$) and comparison between the calculated n -values for ascorbic acid degradation in anaerobic condition ($n = 0.84$) and in aerobic conditions ($n = 0.44$).

Conditions	Temperature (°C)	$k \pm \text{SD}^a$	R^2	R^2 for $n = 0.84$	R^2 for $n = 0.44$
Anaerobic	95	2.73 ± 0.04	0.996	0.996	–
	105	6.19 ± 0.17	0.999	0.999	–
	115	13.71 ± 0.22	0.999	0.999	–
	125	28.75 ± 0.61	0.999	0.999	–
Aerobic	95	552 ± 30	0.994	–	0.994
	105	604 ± 0	0.996	–	0.997
	115	599 ± 39	0.996	–	0.996
	125	418 ± 3	0.990	–	0.987

^a SD = Standard deviation.




Fig. 3. Effect of temperature on the rate constant of ascorbic acid degradation in anaerobic conditions (filled square) referred to the left axis and in aerobic conditions (filled triangle) referred to the right axis.

3.2. Aerobic conditions

In presence of air, ascorbic acid had totally disappeared after 240 min at 95, 105 and 115 °C. All four temperatures gave almost identical degradation curves, except that at 125 °C which had a slight tail (Fig. 4). In aerobic conditions, the temperature has a limited effect on the ascorbic acid degradation in this range. In fact, the degradation curves were superimposed regardless of temperature. We assumed that the degradation observed was due to both the anaerobic mechanism described above and an aerobic component connected with O_2 presence.

Three points called for attention:

A first point was that ascorbic acid degradation was enhanced in the presence of air instead of nitrogen in the same system: even at




Fig. 4. Ascorbic acid degradation in aerobic conditions at 95 °C (◇), 105 °C (□), 115 °C (Δ) and 125 °C (x) in Mcllvaine buffer pH = 3.5.

canning temperatures, the aerobic mechanism is present and predominant. Though O₂ saturation concentrations are low at elevated temperature, the model of Tromans (1998) for O₂ solubility indicates about 16 mg/L at 0.3 MPa. This value was determined using the model established by Tromans and described by equation (8) (Tromans, 1998):

$$[O_2] = P_{O_2} \cdot \exp\left(\frac{0.064 \cdot T^2 + 203.357 \cdot T \cdot \ln\left(\frac{T}{298}\right) - (299.378 + 0.092 \cdot T)(T - 298) - 20.591 \cdot 10^3}{R \cdot T}\right) \quad (8)$$

where [O₂] is the dissolved oxygen concentration in mol.L⁻¹ and P_{O₂} is the oxygen partial pressure in MPa.

In our experiment, the air pressure was set at 0.3 MPa, so the partial pressure of oxygen was P_{O₂} = 0.21 * 0.3 MPa = 0.063 MPa.

The values of dissolved oxygen concentration depending on temperature are reported in Table 2.

There is a non-negligible amount of dissolved oxygen in the heating media, which could explain the importance of the aerobic pathway compared to anaerobic pathway.

A second point was that the four curves were very similar, so the effect of temperature was very limited, while in anaerobic conditions a classical acceleration of degradation rate with temperature was recorded. As it can be hypothesized that this mechanism is still present, purely oxidative degradation would decrease with increasing temperature, which demands further investigation.

A last point is that ascorbic acid degradation did not follow first-order kinetics, nor zero-order kinetics in these conditions.

An *n*-value of 0.44 was obtained by fitting the curves. Therefore, *n* = 0.5 was used to fit the kinetics. To verify if this simplification is coherent, the R² was compared for *n* = 0.44 and *n* = 0.5. A good fit was obtained for all four temperatures using *n* = 0.5 (R² > 0.99) and the R² was the same for the two different *n*-values (Table 1). The values of 0.5-order kinetic constants are reported in Table 1.

The kinetic constant did not vary significantly with temperature and did not fit in Arrhenius plot (Fig. 3). It was impossible to determine an activation energy for aerobic ascorbic acid degradation during sterilization. In aerobic conditions at sterilization temperatures, there was an unidentified factor in ascorbic acid degradation.

4. Conclusion

Ascorbic acid degradation follows first-order kinetics at high temperatures (>100 °C) under anaerobic conditions but it follows 0.5-order kinetics in aerobic conditions. In anaerobic conditions, the good fit of first-order kinetics confirmed that there was only one involved mechanism, the hydrolytic pathway, compared to aerobic conditions where two pathways at least are involved.

Table 2

Concentration of dissolved oxygen in the heating media depending on temperature, according to Tromans's model.

Temperature (°C)	[O ₂] (mg L ⁻¹)
95	15.5
105	15.6
115	15.9
125	16.5

Oxygen is a major factor in ascorbic acid degradation. The significance of oxygen was significantly higher (2.5 times) than process temperature, as 50% degradation of ascorbic acid is observed after 240 min and 100 min thermal treatment at 125 °C in anaerobic and aerobic conditions, respectively. Furthermore, the thermal stability of ascorbic acid in anaerobic conditions is higher signifi-

cantly than known stability in aerobic. Indeed, under anaerobic conditions, 320 min treatment even at 125 °C was not sufficient to degrade all the ascorbic acid initially present.

In aerobic conditions, oxygen availability may be reduced due to the high temperatures that may decrease its solubility, but it stays non negligible. Removing oxygen from the medium appeared to be the better solution to preserve the maximum of vitamins during heat drastic treatments, but it can be difficult to achieve in food industry (matrix complexity, needed of specific devices, etc ...).

References

- Al-Zubaidy, M. M. I., & Khalil, R. A. (2007). Kinetic and prediction studies of ascorbic acid degradation in normal and concentrate local lemon juice during storage. *Food Chemistry*, 101(1), 254–259.
- Berlinet, C., Brat, P., Brillouet, J.-M., & Ducruet, V. (2006). Ascorbic acid, aroma compounds, and browning of orange juices related to PET packaging materials and pH. *Journal of the Science of Food and Agriculture*, 86(13), 2206–2212.
- Blasco, R., Esteve, M. J., Frígola, A., & Rodrigo, C. (2004). Ascorbic acid degradation kinetics in mushrooms in a high-temperature short-time process controlled by a thermoresistometer. *LWT - Food Science and Technology*, 37(2), 171–175.
- Bosch, V., Cilla, A., García-Llatas, G., Gilibert, V., Boix, R., & Alegría, A. (2013). Kinetics of ascorbic acid degradation in fruit-based infant foods during storage. *Journal of Food Engineering*, 116(2), 298–303.
- Condón, S., Arrizubieta, M. J., & Sala, F. J. (1993). Microbial heat resistance determinations by the multipoint system with the thermoresistometer TR-SC Improvement of this methodology. *Journal of Microbiological Methods*, 18(4), 357–366.
- Conesa, R., Andreu, S., Fernández, A., Esnoz, A., & Palop, A. (2009). Nonisothermal heat resistance determinations with the thermoresistometer Mastia. *Journal of Applied Microbiology*, 117(2), 506–513.
- Esteve, M. J., Frígola, A., Martorell, L., & Rodrigo, C. (1998). Kinetics of ascorbic acid degradation in green Asparagus during heat processing. *Journal of Food Protection*, 61(11), 1518–1521.
- Garza, S., Teixidó, J. A., Sanchis, V., Viñas, I., & Condón, S. (1994). Heat resistance of *Saccharomyces cerevisiae* strains isolated from spoiled peach puree. *Food Microbiology*, 23(2), 209–213.
- Hsu, H.-Y., Tsai, Y.-C., Fu, C.-C., & Wu, J. S.-B. (2012). Degradation of ascorbic acid in ethanolic solutions. *Journal of Agricultural and Food Chemistry*, 60(42), 10696–10701.
- Jaffe, G. M. (1984). Ascorbic acid. *Encyclopedia of Chemical Technology*, 24, 8–40.
- Jiang, L., Zheng, H., & Lu, H. (2014). Use of Linear and Weibull functions to model ascorbic acid degradation in Chinese Winter Jujube during postharvest storage in light and dark conditions. *Journal of Food Processing and Preservation*, 38(3), 856–863.
- Lin, S. H., & Galalco, J. (1979). Degradation kinetics of ascorbic acid. *Process Biochemistry*, 14(9), 22–24.
- McIlvaine, T. C. (1921). A buffer solution for colorimetric comparison. *Journal of Biological Chemistry*, 49(1), 183–186.
- Mercali, G. D., Jaeschke, D. P., Tessaro, I. C., & Marczak, L. D. F. (2012). Study of vitamin C degradation in acerola pulp during ohmic and conventional heat treatment. *LWT - Food Science and Technology*, 47(1), 91–95.
- Mercali, G. D., Schwartz, S., Marczak, L. D. F., Tessaro, I. C., & Sastry, S. (2014). Effect of the electric field frequency on ascorbic acid degradation during thermal treatment by ohmic heating. *Journal of Agricultural and Food Chemistry*, 62(25), 5865–5870.
- Oey, I., Verlinde, P., Hendrickx, M., & Van Loey, A. (2006). Temperature and pressure stability of L-ascorbic acid and/or [6s] 5-methyltetrahydrofolic acid: A kinetic study. *European Food Research and Technology*, 223(1), 71–77.
- Palop, A., Marco, A., Raso, J., Sala, F. J., & Condón, S. (1997). Survival of heated

- Bacillus coagulans* spores in a medium acidified with lactic or citric acid. *International Journal of Food Microbiology*, 38(1), 25–30.
- Palop, A., Raso, J., Pagà, R., Condón, S., & Sala, F. J. (1999). Influence of pH on heat resistance of spores of *Bacillus coagulans* in buffer and homogenized foods. *International Journal of Food Microbiology*, 46(3), 243–249.
- Palop, A., Sala, F. J., & Condón, S. (1997). Occurrence of a highly heat-sensitive spore subpopulation of *Bacillus coagulans* STCC 4522 and its conversion to a more heat-stable form. *Applied and Environmental Microbiology*, 63(6), 2246–2251.
- Peng, J., Mah, J.-H., Somavat, R., Mohamed, H., Sastry, S., & Tang, J. (2012). Thermal inactivation kinetics of *Bacillus coagulans* spores in tomato juice. *Journal of Food Protection*, 75(7), 1236–1242.
- Raso, J., Palop, A., Bayarte, M., Condón, S., & Sala, F. J. (1995). Influence of sporulation temperature on the heat resistance of a strain of *Bacillus licheniformis* (Spanish Type Culture Collection 4523). *Food Microbiology*, 12, 357–361.
- Sapei, L., & Hwa, L. (2014). Study on the kinetics of vitamin C degradation in fresh strawberry juices. *Procedia Chemistry*, 9, 62–68.
- Serpen, A., & Gökmen, V. (2007). Reversible degradation kinetics of ascorbic acid under reducing and oxidizing conditions. *Food Chemistry*, 104(2), 721–725.
- Soares, N. F. F., & Hotchkiss, J. H. (1999). Comparative effects of De-aeration and package permeability on ascorbic acid loss in refrigerated orange juice. *Packaging Technology and Science*, 12(3), 111–118.
- Torregrosa, F., Esteve, M. J., Frígola, A., & Cortés, C. (2006). Ascorbic acid stability during refrigerated storage of orange-carrot juice treated by high pulsed electric field and comparison with pasteurized juice. *Journal of Food Engineering*, 73(4), 339–345.
- Tromans, D. (1998). Temperature and pressure dependent solubility of oxygen in water: A thermodynamic analysis. *Hydrometallurgy*, 48(3), 327–342.
- Van Boekel, M. A. J. S. (2009). *Kinetic Modeling of Reactions in Foods*. Boca Raton: Taylor & Francis Group.
- Van den Broeck, I., Ludikhuyze, L., Weemaes, C., Van Loey, A., & Hendrickx, M. (1998). Kinetics for isobaric-isothermal degradation of L-ascorbic acid. *Journal of Agricultural and Food Chemistry*, 46(5), 2001–2006.
- Yuan, J. P., & Chen, F. (1998). Degradation of ascorbic acid in aqueous solution. *Journal of Agricultural and Food Chemistry*, 46(12), 5078–5082.
- Zimmermann, M., Longhi, D. A., Schaffner, D. W., & Aragão, G. M. F. (2014). Predicting *Bacillus coagulans* spores inactivation in tomato pulp under non-isothermal heat treatments. *Journal of Food Science*, 79(5), 935–940.