



Everything you always wanted to know about debiasing quantitative spin density maps ... but were afraid to ask

S. Clerjon, J.-M. Bonny

► To cite this version:

S. Clerjon, J.-M. Bonny. Everything you always wanted to know about debiasing quantitative spin density maps ... but were afraid to ask. MRFood, May 2014, Cesena, Italy. hal-02739040

HAL Id: hal-02739040

<https://hal.inrae.fr/hal-02739040>

Submitted on 2 Jun 2020

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

FOODMR 2014 - Cesena, Italy

Tuesday 20 - Friday 23 May 2014

XII International Conference on the Applications of Magnetic Resonance in Food Science: Defining Food by Magnetic Resonance ABSTRACT SUBMISSION

15 February, 2014

Deadline for abstract submission for oral presentations

28 February, 2014

Deadline for abstract submission for poster presentations

Abstracts have to be sent by e-mail as attached files (Microsoft Word for IBM-compatible) to the secretariat of FoodMR 2014

foodmr2014@unibo.it

Abstracts may include: Tables, Schemes, Figures and References.

To maintain a consistent appearance, all abstracts must be prepared in the following style (**see also the template in the next page**):

- maximum space limit is half page
- maximum usable area is 16.0 cm wide and 12.0 cm high
- times new roman font (symbol font for special characters)
- 14 point boldface font for the title
- 12 point underlined font for presenting Author (12 point normal font for other Authors, if any)
- 12 point *italic font* for affiliations
- affiliations distinguished by superscript numbers if necessary
- one blank line left between address and the beginning of the abstract
- one blank line left between the end of the abstract and references/acknowledgment (if present).

Proposed session (select one):

- ☒ Quantitative NMR
- ☐ Quality and Safety
- ☐ On Line non Invasive NMR
- ☐ Foodomics
- ☐ Multiscale Definition of Food
- ☐ New Developments

Presentation preference: ☒ oral or ☐ poster*

*Please select the option "Oral presentation" to indicate that you would like your abstract to be considered for an oral presentation

Everything you always wanted to know about debiasing quantitative spin density maps ... but were afraid to ask

Sylvie Clerjon and Jean-Marie Bonny, *UR370 QuaPA-INRA, 63122 St-Genès-Champanelle, France*

Mapping spin density, for example based on proton or sodium density, is of great interest for the quantitative analysis of food composition and processing by MRI. Boosting NMR sensitivity, e.g. by using array coils or high static fields, allows decreasing the uncertainty of spin density estimates. However, this improvement is counterbalanced by the bias due to the non-uniformity of emission and reception radiofrequency (RF) fields (respectively $f_{1,em}$ and $f_{1,rec}$). Since both responses are sample-dependent, a quantitative strategy for RF bias correction is to map $f_{1,em}$ in the presence of the sample and then to correct the biases due to both emission and reception, by assuming reciprocity ($f_{1,em} = f_{1,rec}$) [1].

We tested this strategy at 4.7 T on a homogeneous phantom using the same coil for emission and reception. We first mapped $f_{1,em}$ using DAM-SP [2], a flexible method for establishing a compromise between the level of noise propagated onto $f_{1,em}$ maps and the width of the range in which the field can be mapped. Secondly, we analyzed the efficiency of reciprocity-based RF bias correction on images. The results are discussed in terms of bias removal, magnification of propagated noise and loss of reciprocity, which occurs in some areas of the coil volume. While improvements were observed, our results underline further concerns for progressing towards full debiasing of quantitative MRI.

Acknowledgment: The MRI experiments were performed at the MR Platform for biological systems at the INRA Center of Clermont-Ferrand, France.

[1] Hoult D.I., The principle of reciprocity in signal strength calculations: a mathematical guide. *Concepts in Magnetic Resonance* 2000; 12 (4): 173 – 187.

[2] Bouhrara M. and Bonny J.-M. B_1 mapping with selective pulses. *Magn Reson Med.* 2012; 68 (5): 1472 - 1480.