



HAL
open science

Infrared spectroscopy as a rapid tool to assess apricot fruit quality

Sylvie Bureau, Catherine M.G.C. Renard, Jean Marc Audergon

► **To cite this version:**

Sylvie Bureau, Catherine M.G.C. Renard, Jean Marc Audergon. Infrared spectroscopy as a rapid tool to assess apricot fruit quality. 16. International Symposium on Apricot Breeding and Culture, Jun 2015, Shenyang, China. hal-02742207

HAL Id: hal-02742207

<https://hal.inrae.fr/hal-02742207>

Submitted on 3 Jun 2020

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

XVI International Symposium on Apricot Breeding and Culture
and XV Chinese National Symposium on Plum and Apricot

ABSTRACTS



ISHS



CSHS



LAAS

June 29-July 3, 2015, Shenyang, China
Edited by: Weisheng LIU, Shuo LIU, Xiaoxue MA

Infrared Spectroscopy as A Rapid Tool to Assess Apricot Fruit Quality.

Sylvie BUREAU¹ and Catherine M.G.C. RENARD¹

INRA, UMR408 Sécurité et Qualité des Produits d'Origine Végétale
F-84000 Avignon
France

¹Université d'Avignon et des Pays de Vaucluse, UMR408 Sécurité et Qualité des Produits d'Origine Végétale
F-84000 Avignon
France

Zeineb Fakhfackh and Jean-Marc AUDERGON

Unité de Génétique et d'Amélioration des Fruits et Légumes, INRA
Domaine Saint Maurice, BP94, F-84143 Montfavet
France

Keywords: Near and Mid-infrared, model robustness.

Abstract

The infrared spectroscopy is known as a rapid tool to study apricot fruit quality, non-destructively using near-infrared (NIR, 800-2500 nm) range and destructively on fruit puree using mid-infrared (MIR, 4000-650 cm^{-1}) range (Bureau et al., 2009a,b). These techniques appear suitable for the determination of soluble solids content (SSC) and titratable acidity (TA). For example, SSC is determined with a R^2 of 0.84 and a root mean square error of cross-validation (RMSECV) of 8% in NIR. In MIR, its performance of prediction is very good with R^2 of 0.96 and RMSECV of 4%.

Eight apricot varieties, representative of the known apricot diversity for quality traits, have been studied between 2005 and 2011. The objective here was to evaluate the robustness of NIR and MIR techniques and to determine the best strategy: is it more efficient to build models every year or to build global models combining all years?

Compared to the method of reference, infrared spectroscopy has been shown to be fast and easy applicable. This technique can be adapted to routine analysis in apricot industries.