

## From in situ quantitative MRI to modeling the structural changes of meat during heating

Mustapha Bouhrara, Sylvie Clerjon, Jean-Louis Damez, J.-M. Bonny

## ▶ To cite this version:

Mustapha Bouhrara, Sylvie Clerjon, Jean-Louis Damez, J.-M. Bonny. From in situ quantitative MRI to modeling the structural changes of meat during heating. 11. International Conference on the Applications of Magnetic Resonance in Food Science, Jun 2012, Wageningen, Netherlands. 2012, 11th. International Conference on the Applications of Magnetic Resonance in Food Science. hal-02748490

## HAL Id: hal-02748490 https://hal.inrae.fr/hal-02748490

Submitted on 3 Jun 2020

**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers. L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

## From *in-situ* quantitative MRI to modeling the structural changes of meat during heating

M. Bouhrara, S. Clerjon, J.L. Damez and J.M. Bonny

INRA, UR370 Qualité des Produits Animaux, F-63122 Saint Genès Champanelle, France.

Modelling the effects of heating is vital to optimise the quality of cooked meats. The advantage of building a model on the basis of in situ measurements is that it relies on neither reductionist hypotheses by studying an intact sample at the scale of the food consumed, nor on simplified hypotheses by taking into account spatial variations in the sample. We present a novel MRI-compatible device that heats the meat sample via a fluid (20 to 75 °C) and a strategy based on dynamic images acquisition at 4.7T during heating and thermal numerical simulation that simultaneously maps deformation (1), proton density (PD) and temperature (2). The deformation field at temperature  $t_i$  was calculated from the geometrical transformation required to pass from the images showing muscle internal structural markers obtained at  $t_{(i-1)}$  to that obtained at  $t_i$  (1). PD measurement consisted of two spin-echo acquisitions that differed only in their echo times. A fast MRI protocol was used (2) since a long TR was necessary to mitigate  $T_1$ dependence. Also, experimental mapping of  $B_1$  was preferred to numerical simulations in order to unbias PD maps (2, 3), since B<sub>1</sub> distribution depends on unknown and evolving factors. The figure shows: deformation fields (C) derived from  $T_2^*$ -weighted images (A), temperature maps (B) and PD maps (D), at 4 temperatures.  $T_2^*$ -weighted images show internal fiducial markers despite the loss of SNR with temperature. Temperature is almost radial and leads to the spatial variation of both deformation and PD. Deformation increases with temperature and accelerates over 58°C due to sarcoplasmic protein denaturation and collagen contraction. Water is then expelled to the exterior of the sample. Lines (C) and (D) show that the most deformed zones are those where PD is highest. The pressure in the muscle increases due to muscle contractions, causing water to transfer towards the exterior. Moreover, the temperature is higher at the superficial area of the sample, leading to the creation of interfascicular channels that can "trap" the migrating water. Finally average models were constructed from local information provided by maps. These models link deformation, proton density and temperature, and provide new information about the underlying mechanisms of meat deformation and water transfers during heating.



1. Bouhrara, M., Lehallier, B., Clerjon, C., Damez, J.L., Bonny, J.M. (2012) Mapping of muscle deformation during heating: in situ dynamic MRI and nonlinear registration. *Magn. Reson. Imaging* 30, pp. 422

2. Bouhrara M., Clerjon S., Damez J.L., Bonny J.M. (2012). In-situ imaging highlights local structural changes during heating: the case of meat. *J. Agric. Food Chem.*, in press

3. Bouhrara, M., Bonny, J.M. (2012)  $B_1$  mapping with selective pulses. *Magn. Reson. Med.* DOI 10.1002/mrm.24146.