



HAL
open science

A framework for nucleus density quantitative mapping corrected for B1-errors

Sylvie Clerjon, J.-M. Bonny

► **To cite this version:**

Sylvie Clerjon, J.-M. Bonny. A framework for nucleus density quantitative mapping corrected for B1-errors. 13.International Conference on the Applications of Magnetic Resonance in Food Science, Jun 2016, Karlsruhe, Germany. 1 p. hal-02794478

HAL Id: hal-02794478

<https://hal.inrae.fr/hal-02794478>

Submitted on 5 Jun 2020

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

A framework for nucleus density quantitative mapping corrected for B_1 -errors

Jean-Marie Bonny, Sylvie Clerjon,

AgroResonance, UR370 QuaPA - INRA, F-63122 Saint-Genès-Champanelle, France

The quantification of nucleus density (ND) by magnetic resonance imaging (MRI) is of major importance for many applications, and particularly in food science (e.g. moisture or salt content). Nevertheless, ND mapping techniques are underemployed, mainly due to the lack of reliability on the obtained ND estimates. Bias is due to the inhomogeneity in the sample to be imaged of both transmit and receive radiofrequency fields. These B_1 -related inhomogeneities should be taken into account because their effects cumulate, and bias the image intensities in a multiplicative way. The problem is badly conditioned so that even a small B_1 deviation could induce large errors in ND estimation.

Assuming equal spatial distributions of transmit and receive radiofrequency fields, we introduced a generic B_1 correction approach consisting in (i) mapping the transmit B_1 field in the presence of the sample, (ii) inferring the bias field and (iii) correcting the ND map using the calculated bias field. Because acquiring ND and B_1 maps require long acquisition times to prevent T_1 -contamination, we developed a method to reconstruct both maps from the same radiofrequency-spoiled gradient echo MRI data. This approach was supplemented by a theoretical framework which allowed predicting the quality of the corrected ND maps in terms of residual bias and spatially-varying uncertainties. By this way, we can readily optimise the different steps required for ND mapping as well as the accuracy over a wide range of radiofrequency field variations. This global approach dedicated to the high-precision mapping of ND by MRI was validated on various simulated and experimental datasets.