

# Sequential preprocessing through ORThogonalization (SPORT) and its application to near infrared spectroscopy

Jean-Michel Roger, Alessandra Biancolillo, Federico Marini

# ▶ To cite this version:

Jean-Michel Roger, Alessandra Biancolillo, Federico Marini. Sequential preprocessing through OR-Thogonalization (SPORT) and its application to near infrared spectroscopy. Chemometrics and Intelligent Laboratory Systems, 2020, 199, 10.1016/j.chemolab.2020.103975. hal-02963019

# HAL Id: hal-02963019 https://hal.inrae.fr/hal-02963019

Submitted on 1 Jun2022

**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers. L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



Distributed under a Creative Commons Attribution - NonCommercial - NoDerivatives 4.0 International License

Version of Record: https://www.sciencedirect.com/science/article/pii/S0169743919308135 Manuscript\_5b07cc1ac465d65a69ec2f88fe60226b

# **1** Sequential Preprocessing through ORThogonalization (SPORT)



- 3 Jean-Michel Roger<sup>1,2</sup>, Alessandra Biancolillo<sup>3</sup>, Federico Marini<sup>4</sup>
- 4 <sup>1</sup>ITAP, Irstea, Montpellier SupAgro, University of Montpellier, Montpellier, France
- 5 <sup>2</sup>ChemHouse Research Group, Montpellier, France
- 6 <sup>3</sup>University of L'Aquila, Department of Physical and Chemistry science, L'Aquila, Italy.
- <sup>7</sup> <sup>4</sup>University of Rome "La Sapienza", Department of Chemistry, Rome, Italy.

### 8 Abstract

9 In spectroscopy, multivariate calibrations more than often include a pre-processing step to 10 reduce the effect of unwanted (not Y-related) sources of variability. Because there are many 11 types of background noise, there are many pre-treatment methods. It is therefore tedious to 12 select and/or combine the best pre-treatments. This article proposes to combine several pre-13 treatments through the use of sequential and orthogonalized partial least squares (SO-PLS), 14 thus leading to a boosting method. The performances and properties of this new method, 15 called Sequential Preprocessing through ORThogonalization (SPORT), are compared to those of 16 a previously published stacking method. SPORT demonstrates very good calibration 17 performances, but also the ability to make significant pretreatment selections.

# 18 Keywords

Spectroscopy, analytical techniques, pretreatments, ensemble, boosting, multi-block analysis,
 sequential and orthogonalized partial least squares (SO-PLS), sequential preprocessing through
 orthogonalization (SPORT)

## 22 Introduction

23 Spectroscopy and, more generally, analytical chemistry techniques provide multivariate 24 outcomes. These instrumental signals are often collected in order to evaluate one or more 25 properties of a product. This evaluation is generally based on calibrating a chemometric 26 model, e.g., by principal component regression (PCA) or partial least square regression (PLSR). 27 However, in these techniques, model building is based on extracting components which 28 account for a relevant share of the variance in the predictor space. Accordingly, whatever 29 phenomenon (wanted or unwanted) has an impact on the data variance can be included in the 30 predictive model, quite often, in the case of spurious variance, with a detrimental effect. For 31 this reason, the calibration of chemometric models generally includes a pre-treatment step, to 32 reduce the effect of these interfering phenomena. Such interferences are generally due to the 33 variation of physical (e.g., temperature) or chemical (e.g., humidity) influence factors. 34 Variations of these factors induce different effects on the measured spectra: the particle size 35 affects the baselines, the temperature affects the position and relative amplitude of the peaks, the geometric configuration of the measurement induces additive and multiplicative effects, 36 37 etc. The wide variety of effects has led to the development of a large number of pre-treatment 38 methods [1]. The removal of a baseline, whatever its shape, is generally done by calculating 39 suitable approximations, either by a polynomial (Detrend [2]), a low-frequency filter (ALS, Asymmetric Least Squares [3]), and then subtracting it from the measured data. The reduction 40 41 of multiplicative effects is achieved either by logarithmic transformation or by normalization 42 (SNV, Standard Normal Variate [2]), possibly weighted (VSN, Variable Sorting for Normalization [4]) or in comparison with a reference spectrum (MSC, Multiplicative Scatter Correction [5]). 43 44 High frequency noise reduction is achieved by low-pass filtering (SG, Savitsky and Golay [6]). 45 The highlighting of spectral details is done by differentiation, usually through the SG algorithm, 46 which allows the computation of derivatives, without noise magnification. The correction of harmful spaces, which can be observed, for example, when transferring a calibration between
two spectrometers, can be done by an orthogonal projection method (EPO, External
Parameter Orthogonalization [7]; TOP, Transfer by Orthogonal Projection [8]).

50 This raises the problem of choosing a suitable pre-treatment. In [9], different methodologies 51 are examined to select the appropriate pre-treatment. This article concludes that "... how 52 extremely difficult it can be to determine which method - of the vast number of available pre-53 processing methods - can successfully help...". In addition, these pre-treatments are often 54 combined. In many articles, it can be read that baseline reductions are associated with 55 normalizations or even derivatives. These choices are often based on a trial and error 56 procedure, where the combination of different pre-treatments is done sequentially, e.g. SNV 57 followed by a derivative. In this regard, it should be noted that the order according to which 58 pre-treatments are applied may be important [9].

59 Another way to manage the choice and association of pre-processing is to use the so-called 60 "ensemble learning" methods [10]. Ensemble learning is similar to that of "data 61 augmentation", very common in deep learning [11]. Ensemble methods use several learning 62 algorithms to achieve better performance than each algorithm used alone. Several approaches 63 exist, as boosting, model averaging, model combination, buckets of models, stacking. All these methods differ roughly in the way the algorithms are aggregated. For instance, in boosting 64 65 approaches weak learners are trained sequentially on different subset of samples (or, more 66 generally, on different weighting scheme over the training individuals) and, in order for each 67 successive model to be better than the predecessors, along the iterations higher weight (or 68 higher probability of being selected) is given to the most difficult samples. Instead, model 69 stacking, which is the form in which ensemble learning has mostly been used in chemometrics, 70 consists in training a meta-model to output a prediction based on the outcomes of the various 71 individual models, which, differently than in boosting can also be heterogeneous in nature; in 72 its simplest form, predictions from the weak learners can be gathered in a new data matrix, 73 which is used to build a final model. In [12], it is proposed to divide the NIR spectra into 74 intervals, then to stack the PLSR models performed on each interval by combining them by 75 linear regression. In [13], it is proposed to stack different PLSR models using different 76 pretreatments on the same data. Twenty different pre-processing operations, based on first 77 and second derivatives, smoothing, SNV, MSC and their combinations were performed on NIR 78 spectrum sets. A similar procedure, i.e., the exploitation of six PLS models (calculated on data 79 preprocessed by diverse preprocessing approaches) as learners in an ensemble learning 80 modelling algorithm was discussed in [14]. The output of this approach is computed by 81 averaging the predicted values computed by its constituent learners.

In the present article, the concept that different preprocessing strategies applied to the same 82 83 set of spectra could result in a multi-block data, and that, as such, the latter could be 84 processed as a whole through dedicated multi-block strategies, is exploited in a boosting approach. In particular, due to its characteristics, the choice of the multi-block strategy to be 85 86 adopted has fallen on sequential and orthogonalized partial least squares regression (SO-PLS, 87 [15-16]), since it allows the possibility of including/excluding blocks, depending on their 88 relevance, of evaluating the incremental contribution of the different matrices and, up to a 89 certain extent, which blocks carry common and distinctive information. The resulting approach 90 has been called Sequential Preprocessing through ORThogonalization (SPORT) and it will be 91 described in detail in the following sections, together with examples of its application to real 92 world NIR data sets.

## 93 Material and methods

#### 94 **Data sets**

95 The proposed method was tested on three sets of real data:

96 Wheat grain data [17]: The NIR transmission spectra of wheat seeds were measured at 100

- 97 wavelengths and used to calibrate the protein content. The data set contained a calibration set
- 98 of 415 samples and a test set of 108 samples.

99 Meat data [18]: The NIR transmission spectra of fine meat slices were measured at 100

100 wavelengths and used to calibrate the fat content. The data set contained a calibration set of

101 172 samples and a test set of 43 samples.

Tablet Data [19]: Near infrared spectra were collected on 310 tablets (spectral range between 7400 and 10507 cm-1) whose relative active substance contents (% w/w) were available. Data were divided by the Duplex algorithm [20] into a training and a test set of 210 and 100 samples, respectively.

106 Wheat and tablet data sets are freely available for download on the website of the 107 Chemometrics and Analytical Technology group of the Copenhagen University (KU): 108 <u>www.models.life.ku.dk/datasets</u>. The meat dataset is freely available for download on the

109 website of the Carnegie Mellon University, at: <u>http://lib.stat.cmu.edu/datasets/tecator</u>.

110

#### **111 The SPORT method**

As anticipated in the Introduction, the proposed method is based on processing the different data matrices which result by preprocessing a spectral data set by different techniques through a multi-block approach called sequential and orthogonalized partial least squares regression (SO-PLS). As the name suggests, Sequential and Orthogonalized-Partial Least Squares (SO-PLS) is a multi-block regression method where the information is sequentially

117 extracted from the different predictor blocks. Very concisely, considering the simplest multi-

block scenario, i.e., the case of two predictor blocks (X<sub>1</sub> and X<sub>2</sub>) used to estimate a y response,

the algorithm can be summarized by the following steps:

120 1. The **y** response is fitted to  $X_1$  by PLS.

121 
$$\mathbf{y} = \mathbf{X}_1 b_{X_1} + e_1 = T_{X_1} Q_{X_1}^T + e_1$$
 (1)

122 2. X<sub>2</sub> is orthogonalized with respect to the scores extracted from the first PLS regression.

123 
$$X_2^{Orth} = \left[I - T_{X_1} \left(T_{X_1}^T T_{X_1}\right)^{-1} T_{X_1}^T\right] X_2$$
 (2)

124 3. The orthogonalized  $X_2$  is used to predict the y-residuals obtained from step 1.

125 
$$e_1 = X_2^{Orth} b_{X_2^{Orth}} + e_2 = T_{X_2^{Orth}} Q_{X_2^{Orth}}^T + e_2$$
 (3)

4. The full predictive model is calculated summing up results from step 1. and step 3.

127 
$$\hat{y} = X_1 b_{X_1} + X_2^{Orth} b_{X_2^{Orth}} = T_{X_1} Q_{X_1}^T + T_{X_2^{Orth}} Q_{X_2^{Orth}}^T$$
 (4).

128 In equations (1)-(4), b indicate regression coefficients, while T and Q the X-scores and Y-129 loadings, respectively.

130 If more data blocks are involved, they are orthogonalized with respect to all the previous 131 modelled predictors and then used to estimate the **y**-residuals. A wider discussion over the 132 algorithm and the advantages/disadvantages of the method can be found in [15,21]. In SPORT, 133 several pretreatments of the same data block are associated in an SO-PLS.

For each dataset, SPORT was applied with different pretreatments. A repeated cross-validation (2 random blocks by 10 repetitions) was performed for a number of latent variables varying from 0 to 15 for each pretreatment. The results of this cross-validation allowed us to choose the appropriate number of latent variables to keep for each pretreatment, using the global approach described in [22]. Briefly, this consist in building cross-validated SO-PLS models with all the possible combinations of latent variables (within a fixed maximum). The optimal complexity is then defined inspecting the RMSECVs. Then, a SO-PLS model using these numbers was calibrated on the calibration set and applied on the test set. The root mean squared errors of calibration (RMSEC), cross-validation (RMSECV) and prediction (RMSEP) were calculated at each step of the process.

144 In order to compare the results of SPORT with the ones provided by the stacking approach, the 145 same pre-treatments as in [13] were performed on the wheat and meat datasets, as reported 146 in table 1.

147

148 Table 1: list of the pretreatments applied on the datasets wheat and meat. \*: SG-W-O-D means

savitsky and Golay using a W points wide window, an O<sup>th</sup> order polynomial and a D<sup>th</sup> derivative

Pretreatment					
SG-9-3-0*					
SG-9-4-0					
SG-9-3-1					
SG-9-4-1					
SG-9-3-2					
SG-9-4-2					
SNV					

150

151 The order of the blocks may affect the results of SO-PLS, especially as far as the selection of

152 the blocks is concerned. The tablet dataset was used to test the order in which the blocks are

introduced in SPORT. Five pre-treatments, including "raw data" (i.e., just mean centering), first
and second derivatives, SNV and VSN [21] were selected and combined as reported in Table 2.
Not all the possible orders were tested, but only the ones that reversed the orders of raw,

## 156 differentiating and normalization processes.

4	T. H. D. L'. Chandler		· · P · · · · · · · · · · · · · · · ·	1	1.00
157	Table 2: List of boostings of	pretreatments a	ipplied on the tablet -	aataset, with	aijjerent oraers.

Boosting 1	Boosting 2	Boosting 3
raw data	SNV	SG-15-3-2
SG-15-2-1	raw data	SNV
SG-15-3-2	SG-15-3-2	raw data
SNV	VSN, tol 0.0067, Npar 2	VSN, tol 0.0067, Npar 2
VSN, tol 0.0067, Npar 2	SG-15-2-1	SG-15-2-1

158

159 All the calculations discussed in the present paper were run under Matlab (The Mathworks

160 Inc., MA). All the functions can be freely downloaded from:

161 https://www.chem.uniroma1.it/romechemometrics/research/algorithms/

# 162 **Results and discussion**

- 163 Table 3: Results of the single and ensemble models on wheat and meat datasets. <sup>a</sup>: results
- 164 *taken from* [13].

	Wheat			Meat		
Pre-treatment	LVs	RMSEC	RMSEP	LVs	RMSEC	RMSEP
SG-9-3-0	11	0.53	0.71	6	2.97	2.80
SG-9-4-0	10	0.55	0.78	6	2.97	2.80
SG-9-3-1	8	0.55	0.66	11	2.11	2.09

SG-9-4-1	9	0.53	0.72	14	1.89	2.00
SG-9-3-2	6	0.54	0.52	10	1.97	2.08
SG-9-4-2	8	0.52	0.55	8	1.88	2.13
SNV	10	0.54	0.68	4	2.09	2.01
stacked <sup>a</sup>	-	0.50	0.57	-	1.55	1.82
boosted	0,0,4,0,0,0,11	0.47	0.47	0,0,0,0,0,7,7	1.50	1.65

165

166 Table 3 reports the results of the single block models, of the stacked approach and of the 167 SPORT approach, for both wheat and meat datasets. It can be noted that the models built on 168 raw data (after smoothing and mean centering) have poor performances and, in the case of 169 wheat, also require a large number of latent variables. For the wheat dataset, the best pretreatment is second derivative. It allows us to decrease the RMSEP from 0.70 to 0.51, i.e., a 170 171 gain of 50% in the error variance (from 0.49 to 0.25). Moreover, the number of latent variables 172 decreases from 11 to 6. When looking at the application of individual preprocessing strategies, 173 for the meat dataset, the gain is less spectacular. Indeed, the best first-derivative and SNV 174 approximately give the same RMSEP of about 2.00, in comparison to 2.01 for the model built 175 on raw data (after smoothing and mean centering). However, SNV allows us to decrease 176 dramatically the number of latent variables (from 14 to 4). As reported in [13], the stacked 177 model is slightly outperformed by the best single block pretreatments for the wheat dataset, 178 whilst it gives better performances for the meat dataset (RMSEP decreases from 2.01 to 1.82, 179 i.e. the error variance decreases by 20%). The SPORT models give the best performances on 180 both datasets. On the wheat data set, the resulting RMSEP is 0.47, representing a 55% gain in 181 the error variance compared to the raw model. On the meat data set, the RMSEP is 1.65, 182 representing 32% less error variance than the raw model. Compared to the stacked models, 183 the RMSEP decreases from 0.57 to 0.47 and from 1.82 to 1.65 for the wheat and meat 184 datasets, respectively, resulting in a corresponding improvement in error variance of 32 and

185 18%.

186 For both datasets, it can be noted that SPORT led to a very parsimonious selection of blocks, 187 since only two out of 7 blocks were used in the multi-block process. In both cases, a linear 188 pretreatment (SG) was associated with a non-linear pretreatment (SNV). For the wheat 189 dataset, SO-PLS has associated a first derivative (block 3) with SNV (block 7). This may seem 190 surprising at first sight, as the second derivative blocks (5 and 6) give much better performance 191 than the first derivative blocks (3 and 4). This illustrates well the ability of SO-PLS to identify 192 the blocks for additional (non-redundant) information, thanks to the orthogonalization steps 193 that take place between the addition of each new block. For meat data, the same 194 phenomenon can be observed. SO-PLS has associated a second derivative (block 6) with the 195 SNV (block 7), while the first derivative blocks (3 and 4) were the ones giving better 196 performances, when taken individually. In [16], Fig. 4 shows that stacking has also selected a 197 second derivative with SNV. This shows a good agreement between the two methods, with 198 regard to the selection of pre-treatments. Therefore, the reason of the difference in 199 performances observed should be ascribed to something else. It can be hypothesized that 200 stacking aggregation, which consists of a linear combination of the predictions of single block 201 PLS, is less effective than the iterative extraction of complementary information by SO-PLS. In 202 addition to the results obtained on our examples, this hypothesis is based on the fact that in 203 the case where the blocks are made up of different pretreatments of the same initial data set, 204 they contain a large redundancy of information, which is a situation that has to be handled 205 with precaution in a linear regression.

#### 206 Table 4: Results of different boosting orders on the tablet dataset.

block number	Boosting 1	Boosting 2	Boosting 3
1	raw data	SNV	SG-15-3-2

2	SG-15-2-1	raw data	SNV		
3	SG-15-3-2	SG-15-3-2	raw data		
4	SNV	VSN, tol 0.0067, Npar 2	VSN, tol 0.0067, Npar 2		
5	VSN, tol 0.0067, Npar 2	SG-15-2-1	SG-15-2-1		
#LV	0,3,0,0,4	0,5,0,2,0	0,0,5,2,0		
RMSEC	0.27	0.28	0.28		
RMSEP	0.33	0.34	0.34		

207

208 Table 4 shows the influence of the order in which the blocks are introduced into SPORT on the 209 tablet data. First of all, it can be noted that the performances obtained, both in RMSEC and 210 RMSEP, are not very much influenced by the order of the blocks. The total number of latent 211 variables used is the same in any order (7), but is distributed differently over the selected 212 blocks. SO-PLS systematically selected pairs of pre-treatments. It can be noted that the second 213 derivative is never selected. The SNV is not selected either, to the benefit of VSN, which is on 214 the contrary, always selected. This is in line with the fact that, on the one hand, VSN is a 215 generalization of SNV and is therefore supposed to give better results [21] and that, on the 216 other hand, at least one standardization method is selected because the tablet spectra contain 217 a very strong multiplicative effect [1,18]. The "raw data" + VSN pair is selected twice out of 218 three (boostings 2 and 3). In both cases, the block of the first derivative (SG1) was arranged 219 after these two blocks. On the other hand, if block SG1 is interposed between the raw data 220 block and the VSN block, it is selected instead of the raw data block.

### 221 **Conclusion**

This paper explores a new way to associate and select pre-treatments of spectra before their use in a calibration model. This involves applying the different pre-treatments to the same set of spectra, then combining the resulting blocks of data through a multi-block approach (SO- 225 PLS). As it proceeds by a sequential inclusion of the blocks that carry non-redundant 226 information, the proposed method, called sequential preprocessing through orthogonalization 227 (SPORT), belongs to boosting family of ensemble methods. The application of SPORT to 228 different datasets shows the advantage of associating pre-treatments in an ensemble method. 229 It also shows that our boosting approach, on the data sets used, is more efficient than the 230 stacking approach, already published. On the other hand, the influence of the order according 231 to which the individual blocks are processed was also studied. It appears that this order may, 232 as expected, influence which blocks are combined in order to build the final model, but it has a rather small impact on the predictive performances. The use of SPORT in other, more complex 233 234 data sets and applications, such as discrimination, will allow for a more in-depth study of its 235 properties.

#### 236 **References**

237 [1] Zeaiter, M., Rutledge, D.N., in Tauler, R., Walczak, B., & Brown, S. D. (2009). 238 Comprehensive chemometrics: chemical and biochemical data analysis. Elsevier, 121-227 239 [2] Barnes, R. J.; Dhanoa, M. S.; Lister, S. J. (1993). Correction of the Description of 240 Standard Normal Variate (SNV) and De-Trend Transformations in Practical Spectroscopy 241 with Applications in Food and Beverage Analysis, 2nd ed. J. Near Infrared Spectrosc. 1, 242 185-186. 243 [3] Eilers, P. (2004). Parametric time warping. Analytical Chemistry, 76, 404-411. Rabatel, G., Marini, F., Walczak, B., & Roger, J. M. (2019). VSN: Variable sorting for 244 [4] 245 normalization. Journal of Chemometrics, e3164. 246 [5] Isaksson, T.; Næs, T. (1988). The Effect of Multiplicative Scatter Correction (MSC) and 247 Linearity Improvement in NIR Spectroscopy. Appl. Spectrosc. 42, 1273–1284. 248 [6] Savitzky, A.; Golay, M. J. E., (1972). Smoothing and differentiating of data by

simplified least squares procedure. Analytical Chemistry, 11 (44), 1906.

- [7] Roger, J-M.; Chauchard, F.; Bellon-Maurel., V. (2003). EPO–PLS external parameter
   orthogonalisation of PLS. Application to temperature-independent measurement of sugar
   content of intact fruits. *Chemometrics and Intelligent Laboratory Systems*, 66, 191–204.
- [8] Andrew, A.; Fearn, T. (2004). Transfer by orthogonal projection: making near-infrared
  calibrations robust to between-instrument variation. *Chemometrics and Intelligent Laboratory Systems*, 72(1), 51-56.
- [9] Engel, J., Gerretzen, J., Szymańska, E., Jansen, J. J., Downey, G., Blanchet, L., &
  Buydens, L. M. (2013). Breaking with trends in pre-processing?. *TrAC Trends in Analytical Chemistry*, *50*, 96-106.
- [10] Dietterich, T. G. (2000, June). Ensemble methods in machine learning.
  In *International workshop on multiple classifier systems* (pp. 1-15). Springer, Berlin,
  Heidelberg.
- [11] Wong, S. C., Gatt, A., Stamatescu, V., & McDonnell, M. D. (2016, November).
  Understanding data augmentation for classification: when to warp?. In 2016 international *conference on digital image computing: techniques and applications (DICTA)* (pp. 1-6).
  IEEE.
- [12] Ni, W., Brown, S. D., & Man, R. (2009). Stacked partial least squares regression
  analysis for spectral calibration and prediction. *Journal of Chemometrics: A Journal of the Chemometrics Society*, 23(10), 505-517.
- [13] Xu, L., Zhou, Y. P., Tang, L. J., Wu, H. L., Jiang, J. H., Shen, G. L., & Yu, R. Q. (2008).
  Ensemble preprocessing of near-infrared (NIR) spectra for multivariate
  calibration. *Analytica chimica acta*, *616*(2), 138-143.
- 272 [14] Reda R., Saffaj T., Ilham B., Saidi O., Issam K., Brahim L., El Hadrami E. M. (2019). A
  273 comparative study between a new method and other machine learning algorithms for soil

- organic carbon and total nitrogen prediction using near infrared spectroscopy, *Chemometrics and Intelligent Laboratory Systems*, 195, 103873.
- [15] Næs T., Tomic O., Mevik B.H., Martens H. (2011). Path modelling by sequential PLS
  regression. *Journal of Chemometrics: A Journal of the Chemometrics Society.* 25, 28–40.
- 278 [16] Biancolillo, A., & Næs, T. (2019). The Sequential and Orthogonalized PLS Regression
- for Multiblock Regression: Theory, Examples, and Extensions. *Data Handling in Science and Technology*, *31*, 157-177.
- [17] Nielsen, J. P., Pedersen, D. K., & Munck, L. (2003). Development of nondestructive
  screening methods for single kernel characterization of wheat. Cereal chemistry, 80(3),
  274-280.
- [18] Borggaard, C., & Thodberg, H. H. (1992). Optimal minimal neural interpretation of
  spectra. Analytical chemistry, 64(5), 545-551.
- [19] Dyrby, M., Engelsen, S. B., Nørgaard, L., Bruhn, M., & Lundsberg-Nielsen, L. (2002).
  Chemometric quantitation of the active substance (containing C≡ N) in a pharmaceutical
  tablet using near-infrared (NIR) transmittance and NIR FT-Raman spectra. Applied
  Spectroscopy, 56(5), 579-585.
- 290 [20] R.D. Snee (1977). Validation of regression models: methods and examples.
  291 Technometrics, 19 415-428.
- 292 [21] Biancolillo A., Næs T., Bro R., Måge I. (2014). Extension of SO-PLS to multi-way arrays:
- 293 SO-N-PLS, *Chemometrics and Intelligent Laboratory System*, 164, 113–126.
- [22] T. Næs, O. Tomic, B.-H. Mevik, H. Martens. (2011). Path modelling by sequential PLS
  regression, J. Chemometr. 25, 28–40.