

A portable and unilateral nmr device to measure tree water and locate conductive tissues

Shannan Blystone, Guilhem Pagès, Hervé H. Cochard, Pierre Conchon

▶ To cite this version:

Shannan Blystone, Guilhem Pagès, Hervé H. Cochard, Pierre Conchon. A portable and unilateral nmr device to measure tree water and locate conductive tissues. Congrès Italian-French International Conference on Magnetic Resonance, Sep 2022, Milan, Italy. hal-03822545

HAL Id: hal-03822545 https://hal.inrae.fr/hal-03822545

Submitted on 20 Oct 2022

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers. L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

A PORTABLE AND UNILATERAL NMR DEVICE TO MEASURE TREE WATER CONTENT AND LOCATE CONDUCTIVE TISSUES

S. Blystone, † G. Pagès, † H. Cochard †, P. Conchon †

[‡]INRAE, UR QuaPA, F-63122 Saint-Genès Champanelle, France

[‡]INRAE, PROBE research infrastructure, AgroResonance facility, F-63122 Saint-Genès-Champanelle, France

[†]Université Clermont Auvergne, INRAE, PIAF, 63000 Clermont-Ferrand, France

E-mail: shannan.blystone@inrae.fr

Keywords: low field NMR, MRI, instrumentation, plants, trees

The use of NMR technology in the plant sciences has traditionally been limited due to the immobility of the devices, and restrictions with regard to sample size and shape. To overcome these limitations and to be able to study plants directly in their natural environment, we evaluated the NMR capacities of a portable, unilateral magnet: The Nuclear Magnetic Resonance Mobile Universal Surface Explorer (NMR-MOUSE), designed by Blümich et al. [1]. This device permits measuring the NMR signal in increments of up to 100-micrometers. and within a depth of approximately 25-millimeters. We tested the capacity of this device to measure tree water content, and its variation over time, by following the dehydration dynamic of cut branches from six different species and two different functional types. There was a significant linear correlation between the integral of the NMR profiles obtained and the water content of the branches. This significant correlation was present regardless of tree species or functional type. We were also able to visualize the dehydration dynamic of individual branches, over time, through the NMR profiles. We then tested the capacity of the device to differentiate the conductive tissues, i.e. the xylem and phloem fluxes, within conifer branches. The NMR profiles of the branches presented distinct peaks which corresponded to the xylem and phloem tissues, whose location was validated with x-ray microtomography imaging which allows the high-resolution visualization of the tissues within the sample (see Fig. 1). In conclusion, the NMR-MOUSE is a promising candidate for measuring plant water dynamics in the field. Future work will test the capacity of this device to measure tree water content insitu, and to measure the speed of both the xylem and phloem fluxes.




Fig. 1. The graph presents the NMR profile of a silver fir branch, showing the distribution of water within the branch. The peaks correspond to the xylem and phloem tissues. The microtomography image shows the cross-section of the branch, enabling us to precisely locate the depth of the xylem and phloem tissues.

References

[1] B. Blümich, P. Blümler, G. Eidmann, A. Guthausen, R. Haken, U. Schmitz, et al. Magnetic Resonance Imaging 16, 5-6: 479-484 (1998)